

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID: M119744
Date Received: 06/05/08
Date Extracted: 06/10/08
Date Analyzed: 06/10/08
Matrix: Water
Units: ug/L (ppb)

Client: Alaskan Copper Works
Project: PO M119744, F&BI 806067
Lab ID: 806067-01 x10
Data File: 806067-01 x10.031
Instrument: ICPMS1
Operator: hr

Internal Standard:	% Recovery:	Lower Limit:	Upper Limit:
Germanium	96	60	125

Analyte:	Concentration ug/L (ppb)
Chromium	335
Nickel	359
Copper	381
Zinc	18.5

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Analysis For Total Metals By EPA Method 200.8

Client ID:	Method Blank	Client:	Alaskan Copper Works
Date Received:	Not Applicable	Project:	PO M119744, F&BI 806067
Date Extracted:	06/10/08	Lab ID:	I8-216 mb
Date Analyzed:	06/10/08	Data File:	I8-216 mb.016
Matrix:	Water	Instrument:	ICPMS1
Units:	ug/L (ppb)	Operator:	hr

Internal Standard:	% Recovery:	Lower	Upper
Germanium	97	Limit:	Limit:
		60	125

Analyte:	Concentration ug/L (ppb)
Chromium	<1
Nickel	<1
Copper	<1
Zinc	<1

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Date of Report: 06/12/08

Date Received: 06/05/08

Project: Metro Self Monitor, PO M119744, F&BI 806067

**QUALITY ASSURANCE RESULTS
FOR THE ANALYSIS OF WATER SAMPLES
FOR TOTAL METALS USING EPA METHOD 200.8**

Laboratory Code: 806039-02 (Duplicate)

Analyte	Reporting Units	Sample Result	Duplicate Result	Relative Percent Difference	Acceptance Criteria
Chromium	ug/L (ppb)	<1	<1	nm	0-20
Nickel	ug/L (ppb)	9.42	9.97	6	0-20
Copper	ug/L (ppb)	<1	<1	nm	0-20
Zinc	ug/L (ppb)	6.13	6.00	2	0-20

Laboratory Code: 806039-02 (Matrix Spike)

Analyte	Reporting Units	Spike Level	Sample Result	Percent Recovery MS	Acceptance Criteria
Chromium	ug/L (ppb)	20	<1	107	50-150
Nickel	ug/L (ppb)	20	9.42	103 b	50-150
Copper	ug/L (ppb)	20	<1	99	50-150
Zinc	ug/L (ppb)	50	6.13	85	50-150

Laboratory Code: Laboratory Control Sample

Analyte	Reporting Units	Spike Level	Percent Recovery LCS	Acceptance Criteria
Chromium	ug/L (ppb)	20	108	70-130
Nickel	ug/L (ppb)	20	106	70-130
Copper	ug/L (ppb)	20	106	70-130
Zinc	ug/L (ppb)	50	84	70-130

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

Data Qualifiers & Definitions

- a - The analyte was detected at a level less than five times the reporting limit. The RPD results may not provide reliable information on the variability of the analysis.
- A1 - More than one compound of similar molecule structure was identified with equal probability.
- b - The analyte was spiked at a level that was less than five times that present in the sample. Matrix spike recoveries may not be meaningful.
- ca - The calibration results for this range fell outside of acceptance criteria. The value reported is an estimate.
- c - The presence of the analyte indicated may be due to carryover from previous sample injections.
- d - The sample was diluted. Detection limits may be raised due to dilution.
- ds - The sample was diluted. Detection limits are raised due to dilution and surrogate recoveries may not be meaningful.
- dv - Insufficient sample was available to achieve normal reporting limits and limits are raised accordingly.
- fb - The analyte indicated was found in the method blank. The result should be considered an estimate.
- fc - The compound is a common laboratory and field contaminant.
- hr - The sample and duplicate were reextracted and reanalyzed. RPD results were still outside of control limits. The variability is attributed to sample inhomogeneity.
- ht - The sample was extracted outside of holding time. Results should be considered estimates.
- ip - Recovery fell outside of normal control limits. Compounds in the sample matrix interfered with the quantitation of the analyte.
- j - The result is below normal reporting limits. The value reported is an estimate.
- J - The internal standard associated with the analyte is out of control limits. The reported concentration is an estimate.
- jl - The analyte result in the laboratory control sample is out of control limits. The reported concentration should be considered an estimate.
- jr - The rpd result in laboratory control sample associated with the analyte is out of control limits. The reported concentration should be considered an estimate.
- js - The surrogate associated with the analyte is out of control limits. The reported concentration should be considered an estimate.
- lc - The presence of the compound indicated is likely due to laboratory contamination.
- L - The reported concentration was generated from a library search.
- nm - The analyte was not detected in one or more of the duplicate analyses. Therefore, calculation of the RPD is not applicable.
- pc - The sample was received in a container not approved by the method. The value reported should be considered an estimate.
- pr - The sample was received with incorrect preservation. The value reported should be considered an estimate.
- ve - The value reported exceeded the calibration range established for the analyte. The reported concentration should be considered an estimate.
- vo - The value reported fell outside the control limits established for this analyte.
- x - The pattern of peaks present is not indicative of diesel.
- y - The pattern of peaks present is not indicative of motor oil.

Send Report To Glenn Thompson
Company ALASKAN Copper Works
Address 628 S. Aurora St
City, State, ZIP Seattle WA 98134
Phone # 206-571-6033 Fax # 206-382-4309

SAMPLERS (*signature*)

PROJECT NAME/NO.
METRO Self monitor

PO #
119744

REMARKS

Page # _____ of _____

TURNAROUND TIME

☐ Standard (2 Weeks)



☒ RUSH 4 day
Rush charges authorized by:

SAMPLE DISPOSAL

- ☐ Dispose after 30 days
- ☐ Return samples
- ☐ Will call with instructions

[illegible]

Friedman & Bruya, Inc.
3012 16th Avenue West
Seattle, WA 98119-2029
Ph. (206) 285-8282
Fax (206) 283-5044

SIGNATURE	PRINT NAME	COMPANY	DATE	TIME
Relinquished to: 	Gerald Thompson	ACW	5/5/88	2:38 PM
Received by: 	D. A. W.	FBI	"	3:30 PM
Relinquished to:				
Received by:				

FRIEDMAN & BRUYA, INC.

ENVIRONMENTAL CHEMISTS

James E. Bruya, Ph.D.
Charlene Morrow, M.S.
Yelena Aravkina, M.S.
Bradley T. Benson, B.S.
Kurt Johnson, B.S.

3012 16th Avenue West
Seattle, WA 98119-2029
TEL: (206) 285-8282
FAX: (206) 283-5044
e-mail: fbi@isomedia.com

June 12, 2008

Gerry Thompson, Project Manager
Alaskan Copper Works
628 South Hanford
Seattle, WA 98134

Dear Mr. Thompson:

Included are the results from the testing of material submitted on June 5, 2008 from the Metro Self Monitor, PO M119744, F&BI 806067 project. There are 4 pages included in this report. Any samples that may remain are currently scheduled for disposal in 30 days. If you would like us to return your samples or arrange for long term storage at our offices, please contact us as soon as possible.

We appreciate this opportunity to be of service to you and hope you will call if you should have any questions.

Sincerely,

FRIEDMAN & BRUYA, INC.



Michael Erdahl
Project Manager

Enclosures
ACU0612R.DOC